



pH BUFFER SOLUTIONS

All calibration procedures assume that the labeled values of the calibration buffers are correct. But buffer values can change over time and so can your results. A complete range of patented buffer solutions provides pH stability up to 5 years, something never achieved before. The pH buffers 9.21 and 10.01 are even stable in air. High buffering capacity provides rapid, stable calibration. The growth of fungus and micro-organisms is prevented.

Traceability

An important issue for the production of Certified Reference Materials is to ensure traceability through an unbroken chain of comparisons to reference material of the highest metrological quality (Primary Reference Material) from NIST and PTB2. Unlike other manufacturers, where only topdown traceability is applied, Hamilton works with circular or closed-loop traceability, providing unique reliability of Hamilton DuraCal buffers.

Features

- ▶ Convenient 250 mL or 500 mL bottle with built-in calibration compartment
- ▶ Economical, only about 15 mL of buffer is used per calibration
- ▶ Certified pH value from a DAkkS laboratory accredited for pH measurement
- ▶ First class certificate with traceability to international standards
- ▶ Certificates available at www.hamiltoncompany.com
- ▶ Expiration date on the bottle
- ▶ Immune to microbial growth

Top-down traceability: At Hamilton, the pH value of DuraCal buffers is determined by comparison against two secondary reference buffer solutions from accredited suppliers of secondary reference materials. The solutions themselves are compared against primary reference solutions from PTB or NIST. The measurement uncertainties of every measurement comparison are known and documented.

Bottom-up traceability: To ensure the highest possible accuracy and full reliability of the pH value, a representative number of samples from every single production lot is verified by an external, independent and impartial DAkkS3 laboratory. The DuraCal samples are compared against secondary reference solutions from DAkkS and these are referenced themselves to primary reference solutions from PTB or NIST. At this stage, the traceability loop is closed. DAkkS provides Hamilton with a calibration certificate for every DuraCal production batch.

Certified reference material: Due to the complete traceability of the measurement procedure and the assignment of uncertainties to the particular testing steps, the buffers pH 4.01, 7.00, 9.21 and 10.01 are classified as "Certified Reference Material" (CRM).

APPLICATION NOTE
*Standard Solutions
you can trust*

PH BUFFERS

pH Value	Accuracy	Stability*	Certified	By Packaging Unit	PN
1.09	±0.02	60	Hamilton	500 mL	238271
1.68	±0.02	60	Hamilton	500 mL	238272
2.00	±0.02	60	Hamilton	500 mL	238273
3.06	±0.02	60	Hamilton	500 mL	238274
4.01	±0.01/±0.02	24/60	DAkKS	250 mL	238317
4.01	±0.01/±0.02	24/60	DAkKS	500 mL	238217
4.01	±0.01/±0.02	24/60	DAkKS	3 x 500 mL	238917
4.01	±0.01/±0.02	24/60	DAkKS	5 L	238332
4.01	±0.01/±0.02	24/60	DAkKS	10 L	238194
4.01	±0.01/±0.02	24/60	DAkKS	1000 L	238895
5.00	±0.02	60	Hamilton	500 mL	238275
6.00	±0.02	60	Hamilton	500 mL	238276
7.00	±0.01/±0.02	24 / 60	DAkKS	250 mL	238318
7.00	±0.01/±0.02	24 / 60	DAkKS	500 mL	238218
7.00	±0.01/±0.02	24 / 60	DAkKS	3 x 500 mL	238918
7.00	±0.01/±0.02	24 / 60	DAkKS	5 L	238333
7.00	±0.01/±0.02	24 / 60	DAkKS	10 L	238188
7.00	±0.01/±0.02	24 / 60	DAkKS	1000 L	238896
8.00	±0.02	60	Hamilton	500 mL	238277
9.21	±0.02	60	DAkKS	250 mL	238319
9.21	±0.02	60	DAkKS	500 mL	238219
9.21	±0.02	60	DAkKS	3 x 500 mL	238919
9.21	±0.02	60	DAkKS	10 L	238216
9.21	±0.02	60	DAkKS	1000 L	238897
10.01	±0.02	60	DAkKS	250 mL	238321
10.01	±0.02	60	DAkKS	500 mL	238223
10.01	±0.02	60	DAkKS	3 x 500 mL	238923
10.01	±0.02	60	DAkKS	10 L	238187
10.01	±0.02	60	DAkKS	1000 L	238898
11.00	±0.02	24	Hamilton	500 mL	238278
12.00	±0.02	24	Hamilton	500 mL	238279
4.01/7.00/9.21	±0.01/±0.02	24/60	DAkKS	500 mL, mixed	238922
4.01/7.00/10.01	±0.01/±0.02	24/60	DAkKS	500 mL, mixed	238924

ORP BUFFERS

pH Value	Accuracy	Stability*	Certified	By Packaging Unit	PN
271 mV	±5 mV	24	None	500 mL	238228
475 mV	±5 mV	24	None	250 mL	238322
475 mV	±5 mV	24	None	500 mL	238227

Simple handling for professional results

Step 1 Open bottle



Step 2 Fill calibration compartment



Step 3 Calibrate electrode



Step 4 Empty calibration compartment

